

Surrogate Summary

SDG No.: **Q1514**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO109425.D	PIBLK-PO109425.D	Tetrachloro-m-xylene	1	20	25.3	126		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	25.3	126		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	24.1	121		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	25.5	127		70 (60)	130 (140)
I.BLK-PO109727.D	PIBLK-PO109727.D	Tetrachloro-m-xylene	1	20	20.5	103		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.1	85		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.2	106		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	18.8	94		70 (60)	130 (140)
PB167028BL	PB167028BL	Tetrachloro-m-xylene	1	20	19.7	99		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	16.8	84		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	20.1	101		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	18.5	92		30 (10)	150 (173)
PB167028BS	PB167028BS	Tetrachloro-m-xylene	1	20	19.0	95		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	17.2	86		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	19.2	96		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	18.9	95		30 (10)	150 (173)
PB167028BSD	PB167028BSD	Tetrachloro-m-xylene	1	20	18.8	94		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	16.9	84		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	19.0	95		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	18.5	92		30 (10)	150 (173)
Q1514-07	ENV-105-GW01	Tetrachloro-m-xylene	1	20	23.7	119		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	13.5	68		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	23.7	119		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	14.7	74		30 (10)	150 (173)
Q1514-08	ENV-103-GW01	Tetrachloro-m-xylene	1	20	23.9	119		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	14.0	70		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	24.4	122		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	15.5	78		30 (10)	150 (173)
Q1514-09	FB03062025	Tetrachloro-m-xylene	1	20	22.9	115		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	15.2	76		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	23.2	116		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	16.5	83		30 (10)	150 (173)
I.BLK-PO109741.D	PIBLK-PO109741.D	Tetrachloro-m-xylene	1	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.3	91		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.2	106		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.0	100		70 (60)	130 (140)
I.BLK-PP069995.D	PIBLK-PP069995.D	Tetrachloro-m-xylene	1	20	21.9	109		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.8	109		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.6	108		70 (60)	130 (140)
I.BLK-PP070348.D	PIBLK-PP070348.D	Tetrachloro-m-xylene	1	20	20.8	104		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: **Q1514**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070348.D	PIBLK-PP070348.D	Decachlorobiphenyl	1	20	19.6	98		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.6	103		70 (60)	130 (140)
PB167029BL	PB167029BL	Decachlorobiphenyl	2	20	20.6	103		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	24.2	121		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.4	107		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	24.0	120		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.6	108		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	24.6	123		30 (32)	150 (144)
PB167029BS	PB167029BS	Decachlorobiphenyl	1	20	22.2	111		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	22.8	114		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.1	110		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	21.4	107		70 (60)	130 (140)
I.BLK-PP070363.D	PIBLK-PP070363.D	Decachlorobiphenyl	1	20	19.8	99		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.8	104		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.3	101		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	13.3	67		30 (32)	150 (144)
Q1508-01MS	RBR251372MS	Decachlorobiphenyl	1	20	13.9	70		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	12.2	61		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.7	74		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	13.8	69		30 (32)	150 (144)
Q1508-01MSD	RBR251372MSD	Decachlorobiphenyl	1	20	14.2	71		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	12.5	62		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.6	73		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	17.6	88		30 (32)	150 (144)
Q1514-01	ENV-105-SB01	Decachlorobiphenyl	1	20	11.3	57		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.1	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	10.8	54		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	22.3	112		30 (32)	150 (144)
Q1514-03	ENV-105-SB02	Decachlorobiphenyl	1	20	13.1	66		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	24.2	121		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	13.9	70		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	21.6	108		30 (32)	150 (144)
Q1514-05	ENV-103-SB01	Decachlorobiphenyl	1	20	15.2	76		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	22.8	114		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	16.3	81		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	20.8	104		70 (60)	130 (140)
I.BLK-PP070378.D	PIBLK-PP070378.D	Decachlorobiphenyl	1	20	18.9	95		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.3	97		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.3	101		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	20.8	104		70 (60)	130 (140)