

Report of Analysis

Client:	Portal Partners Tr	i-Venture			Date Collected:			
Project:	Project: Amtrak Sawtooth Bridges 2025				Date Received:			
Client Sample ID:	PB167090BS				SDG No.:	Q1514		
Lab Sample ID:	PB167090BS				Matrix:	TCLP		
Analytical Method	: SW8151A				% Solid:	0	Decanted:	
Sample Wt/Vol:	1000 Units	mL			Final Vol:	10000	uL	
Soil Aliquot Vol:		uL			Test:	TCLP Herbic	ide	
Extraction Type:					Injection Volume :			
GPC Factor :	1.0	PH :						
Prep Method :	SW3510C							
File ID/Qc Batch:	Dilution:	Prep Date			Date Analyzed	Prep	Prep Batch ID	
PS029398.D	1	03/11/25 11:30			03/12/25 10:21	PB16	PB167090	
CAS Number	Parameter	Conc.	Qualifier	MDL		LOQ / CF	RQL	Units
TARGETS								
94-75-7	2,4-D	4.80		0.92		2	.00	ug/L
93-72-1	2,4,5-TP (Silvex)	5.10		0.78		2	.00	ug/L
SURROGATES 19719-28-9	2,4-DCAA	521		70 (39) -	130 (175)	1	04%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound
- * = Values outside of QC limits
- D = Dilution

 $\mathbf{S}=\mathbf{Indicates}$ estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit