

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 09:04 Initial Calibration Time(s): 10:35 11:29

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.05       | 9.06      | 8.96                   | 9.16 | 0.01       |
| Tetrachloro-m-xylene | 3.54       | 3.54      | 3.44                   | 3.64 | 0.00       |
| gamma-BHC (Lindane)  | 4.33       | 4.33      | 4.23                   | 4.43 | 0.00       |
| Heptachlor           | 4.91       | 4.92      | 4.82                   | 5.02 | 0.01       |
| Heptachlor epoxide   | 5.68       | 5.68      | 5.58                   | 5.78 | 0.00       |
| Endrin               | 6.57       | 6.57      | 6.47                   | 6.67 | 0.00       |
| Methoxychlor         | 7.50       | 7.50      | 7.40                   | 7.60 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 09:04 Initial Calibration Time(s): 10:35 11:29

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Decachlorobiphenyl   | 7.91       | 7.91      | 7.81      | 8.01 | 0.00       |
| Tetrachloro-m-xylene | 2.77       | 2.77      | 2.67      | 2.87 | 0.00       |
| gamma-BHC (Lindane)  | 3.60       | 3.60      | 3.50      | 3.70 | 0.00       |
| Heptachlor           | 3.94       | 3.94      | 3.84      | 4.04 | 0.00       |
| Heptachlor epoxide   | 4.72       | 4.73      | 4.63      | 4.83 | 0.01       |
| Endrin               | 5.63       | 5.63      | 5.53      | 5.73 | 0.00       |
| Methoxychlor         | 6.61       | 6.61      | 6.51      | 6.71 | 0.00       |