

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094664.D Time Analyzed: 22:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.050	8.956	9.156	57.790	50.000	15.6
Endrin	6.570	6.474	6.674	47.710	50.000	-4.6
gamma-BHC (Lindane)	4.325	4.227	4.427	53.770	50.000	7.5
Heptachlor	4.912	4.815	5.015	50.100	50.000	0.2
Heptachlor epoxide	5.680	5.583	5.783	52.710	50.000	5.4
Methoxychlor	7.496	7.400	7.600	52.160	50.000	4.3
Tetrachloro-m-xylene	3.536	3.438	3.638	53.830	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094664.D Time Analyzed: 22:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.904	7.807	8.007	57.440	50.000	14.9
Endrin	5.631	5.534	5.734	52.440	50.000	4.9
gamma-BHC (Lindane)	3.603	3.504	3.704	56.730	50.000	13.5
Heptachlor	3.940	3.842	4.042	52.860	50.000	5.7
Heptachlor epoxide	4.722	4.625	4.825	55.720	50.000	11.4
Methoxychlor	6.604	6.507	6.707	50.980	50.000	2.0
Tetrachloro-m-xylene	2.771	2.672	2.872	56.680	50.000	13.4