

Surrogate Summary

SDG No.: **Q1514**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094566.D	PIBLK-PL094566.D	Decachlorobiphenyl	1	20	22.7	114		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.6	103		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.0	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.2	101		30 (77)	150 (126)
I.BLK-PL094639.D	PIBLK-PL094639.D	Decachlorobiphenyl	1	20	24.1	121		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	21.0	105		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	23.8	119		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.0	105		30 (77)	150 (126)
Q1488-02MS	ENV-101-SB01MS	Decachlorobiphenyl	1	20	27.9	139		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.6	103		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	26.8	134		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.6	108		30 (77)	150 (126)
Q1488-02MSD	ENV-101-SB01MSD	Decachlorobiphenyl	1	20	27.8	139		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.9	104		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	26.7	133		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.8	109		30 (77)	150 (126)
I.BLK-PL094651.D	PIBLK-PL094651.D	Decachlorobiphenyl	1	20	26.8	134		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	22.0	110		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	24.6	123		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	22.4	112		30 (77)	150 (126)
Q1514-02	ENV-105-SB01	Decachlorobiphenyl	1	20	24.5	122		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.6	98		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	23.2	116		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (77)	150 (126)
Q1514-04	ENV-105-SB02	Decachlorobiphenyl	1	20	22.4	112		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.8	104		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.8	109		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.3	106		30 (77)	150 (126)
Q1514-06	ENV-103-SB01	Decachlorobiphenyl	1	20	25.3	127		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.7	103		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	23.7	119		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.9	104		30 (77)	150 (126)
I.BLK-PL094663.D	PIBLK-PL094663.D	Decachlorobiphenyl	1	20	28.3	141		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	23.2	116		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	26.4	132		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	23.4	117		30 (77)	150 (126)
I.BLK-PL094676.D	PIBLK-PL094676.D	Decachlorobiphenyl	1	20	23.1	115		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.2	91		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.0	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.9	94		30 (77)	150 (126)
PB167091BL	PB167091BL	Decachlorobiphenyl	1	20	25.8	129		30 (43)	150 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q1514

Client: Portal Partners Tri-Venture

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
PB167091BL	PB167091BL	Tetrachloro-m-xylene	1	20	21.2	106		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.2	106		30 (43)	150 (140)
PB167091BS	PB167091BS	Tetrachloro-m-xylene	2	20	21.9	109		30 (77)	150 (126)
		Decachlorobiphenyl	1	20	24.4	122		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.2	96		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	20.0	100		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.8	104		30 (77)	150 (126)
PB167020TB	PB167020TB	Decachlorobiphenyl	1	20	21.7	108		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	17.1	86		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	17.9	90		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	17.6	88		30 (77)	150 (126)
I.BLK-PL094694.D	PIBLK-PL094694.D	Decachlorobiphenyl	1	20	24.5	123		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.3	97		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	22.5	112		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.2	96		30 (77)	150 (126)