

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1514 SDG NO.: Q1514

Lab File ID: BG064044.D

DFTPP Injection Date: 03/05/2025

Instrument ID: BNA_G

DFTPP Injection Time: 08:15

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	55.3
68	Less than 2.0% of mass 69	0.4 (1.1) 1
69	Mass 69 relative abundance	38
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	47
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	4.5
441	Present, but less than mass 443	13.7
442	Greater than 50% of mass 198	100.0
443	15.0 - 24.0% of mass 442	17.2 (20.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BG064045.D	03/05/2025	09:02
SSTDICC005	SSTDICC005	BG064046.D	03/05/2025	09:42
SSTDICC010	SSTDICC010	BG064047.D	03/05/2025	10:22
SSTDICC020	SSTDICC020	BG064048.D	03/05/2025	11:03
SSTDICCC040	SSTDICCC040	BG064049.D	03/05/2025	11:43
SSTDICC050	SSTDICC050	BG064050.D	03/05/2025	12:23
SSTDICC060	SSTDICC060	BG064051.D	03/05/2025	13:04
SSTDICC080	SSTDICC080	BG064052.D	03/05/2025	13:44