

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	03/04/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	03/04/25
Client Sample ID:	ENV-101-SB01MS	SDG No.:	Q1514
Lab Sample ID:	Q1488-02MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141893.D	1	03/07/25 12:45	03/07/25 18:30	PB167035

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	290		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	380		8.40	50.0	ug/L
95-48-7	2-Methylphenol	400		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	410		11.5	100	ug/L
67-72-1	Hexachloroethane	370		10.1	50.0	ug/L
98-95-3	Nitrobenzene	450		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	420		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	470		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	500		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	560		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	500		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	950	E	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	118		15 (10) - 110 (139)	79%	SPK: 150
13127-88-3	Phenol-d6	108		15 (10) - 110 (134)	72%	SPK: 150
4165-60-0	Nitrobenzene-d5	110		30 (49) - 130 (133)	110%	SPK: 100
321-60-8	2-Fluorobiphenyl	102		30 (52) - 130 (132)	102%	SPK: 100
118-79-6	2,4,6-Tribromophenol	182	*	15 (44) - 110 (137)	122%	SPK: 150
1718-51-0	Terphenyl-d14	83.1		30 (48) - 130 (125)	83%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	147000		6.887		
1146-65-2	Naphthalene-d8	566000		8.169		
15067-26-2	Acenaphthene-d10	296000		9.922		
1517-22-2	Phenanthrene-d10	482000		11.404		
1719-03-5	Chrysene-d12	376000		14.045		
1520-96-3	Perylene-d12	373000		15.515		

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	03/04/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	03/04/25
Client Sample ID:	ENV-101-SB01MS	SDG No.:	Q1514
Lab Sample ID:	Q1488-02MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141893.D	1	03/07/25 12:45	03/07/25 18:30	PB167035

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products