

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/07/2025

Lab File ID: BF141882.D Time Analyzed: 08:56

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	205331	6.887	782990	8.17	424782	9.92
UPPER LIMIT	410662	7.387	1565980	8.669	849564	10.422
LOWER LIMIT	102666	6.387	391495	7.669	212391	9.422
EPA SAMPLE NO.						
01 ENV-101-SB01MS	146872	6.89	566046	8.17	296352	9.92
02 ENV-101-SB01MSD	134345	6.89	510741	8.17	263683	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.