

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1514 SDG NO.: Q1514

Lab File ID: BG064072.D

DFTPP Injection Date: 03/10/2025

Instrument ID: BNA_G

DFTPP Injection Time: 08:43

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	59.5
68	Less than 2.0% of mass 69	0.5 (1.3) 1
69	Mass 69 relative abundance	40.6
70	Less than 2.0% of mass 69	0.1 (0.3) 1
127	10.0 - 80.0% of mass 198	48.3
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
275	10.0 - 60.0% of mass 198	28
365	Greater than 1% of mass 198	5.1
441	Present, but less than mass 443	14.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.5 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064073.D	03/10/2025	09:23
PB167035BL	PB167035BL	BG064074.D	03/10/2025	10:03
PB167035BS	PB167035BS	BG064075.D	03/10/2025	10:44
PB167020TB	PB167020TB	BG064076.D	03/10/2025	11:24
ENV-105-SB01	Q1514-02	BG064082.D	03/10/2025	15:44
ENV-105-SB02	Q1514-04	BG064083.D	03/10/2025	16:25
ENV-103-SB01	Q1514-06	BG064084.D	03/10/2025	17:05