

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514
 Lab File ID: VN085928.D Date Analyzed: 03/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193697	8.22	310218	9.10	283255	11.87
UPPER LIMIT	387394	8.724	620436	9.6	566510	12.365
LOWER LIMIT	96848.5	7.724	155109	8.6	141628	11.365
EPA SAMPLE NO.						
ENV-105-SB01	154053	8.22	299260	9.10	282311	11.87
ENV-105-SB02	150437	8.22	296435	9.10	287045	11.87
ENV-103-SB01	159259	8.22	315656	9.10	300142	11.87
VN0311WBL01	166099	8.22	319285	9.10	277562	11.87
VN0311WBS01	171959	8.22	282309	9.10	250243	11.87
VN0311WBSD01	166815	8.22	269902	9.10	244088	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.