

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514
 Lab File ID: VN085928.D Date Analyzed: 03/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	153775	13.788				
UPPER LIMIT	307550	14.288				
LOWER LIMIT	76887.5	13.288				
EPA SAMPLE NO.						
ENV-105-SB01	99543	13.79				
ENV-105-SB02	110049	13.79				
ENV-103-SB01	110799	13.79				
VN0311WBL01	106492	13.79				
VN0311WBS01	128349	13.79				
VN0311WBSD01	123338	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.