

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514
 Lab File ID: VN085914.D Date Analyzed: 03/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:21
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	199203	8.22	319135	9.10	288321	11.87
UPPER LIMIT	398406	8.724	638270	9.6	576642	12.365
LOWER LIMIT	99601.5	7.724	159568	8.6	144161	11.365
EPA SAMPLE NO.						
ENV-105-GW01	180602	8.22	336697	9.10	290159	11.87
ENV-103-GW01	199177	8.22	373736	9.10	326066	11.87
TB03062025	185446	8.22	343862	9.10	305300	11.87
VN0310WBL01	185188	8.22	337149	9.10	297684	11.87
VN0310WBS01	200845	8.22	334623	9.10	292041	11.87
VN0310WBSD01	183155	8.22	304090	9.10	267080	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.