

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514
 Lab File ID: VN085928.D Date Analyzed: 03/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193697	8.22	310218	9.10	283255	11.87
UPPER LIMIT	387394	8.724	620436	9.6	566510	12.365
LOWER LIMIT	96848.5	7.724	155109	8.6	141628	11.365
EPA SAMPLE NO.						
FB03062025	176702	8.22	336260	9.10	299219	11.87
VN0311WBL01	166099	8.22	319285	9.10	277562	11.87
VN0311WBS01	171959	8.22	282309	9.10	250243	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.