

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514
 Lab File ID: VN085914.D Date Analyzed: 03/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:21
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	156195	13.788				
UPPER LIMIT	312390	14.288				
LOWER LIMIT	78097.5	13.288				
EPA SAMPLE NO.						
ENV-105-GW01	119785	13.79				
ENV-103-GW01	130206	13.79				
TB03062025	121043	13.79				
VN0310WBL01	112686	13.79				
VN0310WBS01	146257	13.79				
VN0310WBSD01	130620	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.