

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1514 SAS No.: Q1514 SDG NO.: Q1514
 Lab File ID: VY021474.D Date Analyzed: 03/11/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:42
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	170654	13.346				
UPPER LIMIT	341308	13.846				
LOWER LIMIT	85327	12.846				
EPA SAMPLE NO.						
ENV-105-SB01	159534	13.35				
ENV-105-SB02	159684	13.35				
ENV-103-SB01	126386	13.35				
VY0311SBL01	157191	13.35				
VY0311SBS01	162394	13.35				
VY0311SBSD01	150821	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.