



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1519

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134926

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: COD	CCV1	mg/L	50.319	50	101	95-105	03/10/2025
Sample ID: COD	CCV2	mg/L	48.339	50	97	95-105	03/10/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1519

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB134932

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.399	0.4	100	90-110	03/06/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.409	0.4	102	90-110	03/06/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.409	0.4	102	90-110	03/06/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1519

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB135012

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	03/12/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	03/12/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.97	1	97	90-110	03/12/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.95	1	95	90-110	03/12/2025