

Laboratory Control Sample Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1519

Project: Rotor Clip NJ WTD - 2025

Run No.: LB134926

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134926BS							
COD	mg/L	50	49.3		99	1	90-110	03/10/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1519
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB134930

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134930BS							
BOD5	mg/L	198	184		93	1	84.6-115.4	03/07/2025

Laboratory Control Sample Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1519

Project: Rotor Clip NJ WTD - 2025

Run No.: LB134932

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134932BS							
Residual Chlorine	mg/L	0.4	0.41		102	1	90-110	03/06/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1519
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB134949

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134949BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	03/07/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1519
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB134975

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134975BS							
TSS	mg/L	550	532		97	1	90-110	03/10/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1519
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB135012

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167063BS							
Ammonia as N	mg/L	1	0.98		98	1	90-110	03/12/2025