

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1519
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1519 **SAS No.:** Q1519
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	541	520	104	95 - 105	P	03/12/2025	10:52	LB135011
	Copper	532	510	104	95 - 105	P	03/12/2025	10:52	LB135011
	Nickel	517	530	98	95 - 105	P	03/12/2025	10:52	LB135011
	Zinc	1050	1000	105	95 - 105	P	03/12/2025	10:52	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1519
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1519 **SAS No.:** Q1519
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	10.6	10.0	106	80 - 120	P	03/12/2025	10:56	LB135011
	Copper	22.9	20.0	114	80 - 120	P	03/12/2025	10:56	LB135011
	Nickel	38.7	40.0	97	80 - 120	P	03/12/2025	10:56	LB135011
	Zinc	42.6	40.0	106	80 - 120	P	03/12/2025	10:56	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1519

Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1519 **SAS No.:** Q1519

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	1050	1000	105	90 - 110	P	03/12/2025	11:29	LB135011
	Copper	1260	1250	101	90 - 110	P	03/12/2025	11:29	LB135011
	Nickel	2450	2500	98	90 - 110	P	03/12/2025	11:29	LB135011
	Zinc	2530	2500	101	90 - 110	P	03/12/2025	11:29	LB135011
CCV02	Chromium	1070	1000	107	90 - 110	P	03/12/2025	12:20	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	12:20	LB135011
	Nickel	2460	2500	98	90 - 110	P	03/12/2025	12:20	LB135011
	Zinc	2600	2500	104	90 - 110	P	03/12/2025	12:20	LB135011
CCV03	Chromium	1060	1000	106	90 - 110	P	03/12/2025	13:10	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	13:10	LB135011
	Nickel	2480	2500	99	90 - 110	P	03/12/2025	13:10	LB135011
	Zinc	2600	2500	104	90 - 110	P	03/12/2025	13:10	LB135011
CCV04	Chromium	1060	1000	106	90 - 110	P	03/12/2025	14:24	LB135011
	Copper	1290	1250	103	90 - 110	P	03/12/2025	14:24	LB135011
	Nickel	2490	2500	100	90 - 110	P	03/12/2025	14:24	LB135011
	Zinc	2590	2500	103	90 - 110	P	03/12/2025	14:24	LB135011
CCV05	Chromium	1050	1000	105	90 - 110	P	03/12/2025	15:25	LB135011
	Copper	1260	1250	101	90 - 110	P	03/12/2025	15:25	LB135011
	Nickel	2440	2500	98	90 - 110	P	03/12/2025	15:25	LB135011
	Zinc	2530	2500	101	90 - 110	P	03/12/2025	15:25	LB135011
CCV06	Chromium	1050	1000	105	90 - 110	P	03/12/2025	16:25	LB135011
	Copper	1260	1250	100	90 - 110	P	03/12/2025	16:25	LB135011
	Nickel	2430	2500	97	90 - 110	P	03/12/2025	16:25	LB135011
	Zinc	2530	2500	101	90 - 110	P	03/12/2025	16:25	LB135011
CCV07	Chromium	1050	1000	105	90 - 110	P	03/12/2025	17:25	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	17:25	LB135011
	Nickel	2450	2500	98	90 - 110	P	03/12/2025	17:25	LB135011
	Zinc	2560	2500	102	90 - 110	P	03/12/2025	17:25	LB135011
CCV08	Chromium	1070	1000	107	90 - 110	P	03/12/2025	18:23	LB135011
	Copper	1270	1250	102	90 - 110	P	03/12/2025	18:23	LB135011
	Nickel	2470	2500	99	90 - 110	P	03/12/2025	18:23	LB135011
	Zinc	2580	2500	103	90 - 110	P	03/12/2025	18:23	LB135011
CCV09	Chromium	1050	1000	105	90 - 110	P	03/12/2025	18:49	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	18:49	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1519
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1519 **SAS No.:** Q1519
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV09	Nickel	2480	2500	99	90 - 110	P	03/12/2025	18:49	LB135011
	Zinc	2560	2500	102	90 - 110	P	03/12/2025	18:49	LB135011