



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1519

Contract: VERI01

Lab Code: CHEM

Case No.: Q1519

SAS No.: Q1519

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	10.3	10.0	103	40 - 160	P	03/12/2025	11:05	LB135011
	Copper	22.6	20.0	113	40 - 160	P	03/12/2025	11:05	LB135011
	Nickel	38.9	40.0	97	40 - 160	P	03/12/2025	11:05	LB135011
	Zinc	42.3	40.0	106	40 - 160	P	03/12/2025	11:05	LB135011