

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1519
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1519 **SAS No.:** Q1519
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	60.0	52.0	115	42	62	03/12/2025	11:09	LB135011
	Copper	7.26	2.0	363	-18	22	03/12/2025	11:09	LB135011
	Nickel	1.62	2.0	81	-38	42	03/12/2025	11:09	LB135011
	Zinc	6.95			-40	40	03/12/2025	11:09	LB135011
ICSAB01	Chromium	595	542	110	460	624	03/12/2025	11:14	LB135011
	Copper	509	511	100	434	588	03/12/2025	11:14	LB135011
	Nickel	1000	954	105	810	1100	03/12/2025	11:14	LB135011
	Zinc	1080	952	113	809	1095	03/12/2025	11:14	LB135011
ICSA	Chromium	63.4	52.0	122	42	62	03/12/2025	11:21	LB135011
	Copper	10.5	2.0	527	-18	22	03/12/2025	11:21	LB135011
	Nickel	0.050	2.0	2	-38	42	03/12/2025	11:21	LB135011
	Zinc	16.0			-40	40	03/12/2025	11:21	LB135011
ICSAB	Chromium	618	542	114	460	624	03/12/2025	11:25	LB135011
	Copper	553	511	108	434	588	03/12/2025	11:25	LB135011
	Nickel	1030	954	108	810	1100	03/12/2025	11:25	LB135011
	Zinc	1120	952	118	809	1095	03/12/2025	11:25	LB135011