

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1519
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1519 **SAS No.:** Q1519
Matrix: Water **Sample ID:** Q1522-02 **Client ID:** TW-WTS-04DUP
Percent Solids for Sample: NA **Duplicate ID** Q1522-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	5.00	U	5.00	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Nickel	ug/L	20	12.8	J	12.7	J	1		P
Zinc	ug/L	20	15.5	J	18.5	J	18		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1519
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1519 **SAS No.:** Q1519
Matrix: Water **Sample ID:** Q1522-02MS **Client ID:** TW-WTS-04MSD
Percent Solids for Sample: NA **Duplicate ID** Q1522-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	207		205		1		P
Copper	ug/L	20	149		149		0		P
Nickel	ug/L	20	251		250		0		P
Zinc	ug/L	20	119		120		1		P