

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1519

Contract: VERI01

Lab Code: CHEM

Case No.: Q1519

SAS No.: Q1519

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167038BS							
Chromium	ug/L	200	209		104	80 - 120	P
Copper	ug/L	150	152		101	80 - 120	P
Nickel	ug/L	250	235		94	80 - 120	P
Zinc	ug/L	100	105		105	80 - 120	P