



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **ENTACT**

**Project Name: 540 Degraw St, Brooklyn, NY - E9309**

**Project # N/A**

**Chemtech Project # Q1522**

**Test Name: pH,Anions Group2,Flash Point,BOD5,TSS**

### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 03/07/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Anions Group2, BOD5, Flash Point, Hexavalent Chromium, Mercury, Metals Group4, PCB, pH, SVOCMS Group4, TKN, Total Nitrogen, TS, TSS and VOCMS Group4. This data package contains results for pH,Anions Group2,Flash Point,BOD5,TSS.

### **C. Analytical Techniques:**

The analysis of Flash Point was based on method 1010B, The analysis of Anions Group2 was based on method 300.0, The analysis of pH was based on method 9040C, The analysis of TSS was based on method SM2540 D and The analysis of BOD5 was based on method SM5210 B.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for TW-WTS-03 of pH Due to sample was receive out of holding time.

Sample TW-WTS-04 was diluted due to high concentrations for Chloride & Sample TW-WTS-04DL was diluted due to high concentrations for Chloride.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (TW-WTS-04MS) analysis met criteria for all samples except for Chloride due to Sample matrix interferences.

The Matrix Spike Duplicate (TW-WTS-04MSD) analysis met criteria for all samples except for Chloride due to Sample matrix interferences.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

---



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_