

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134955

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/10/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	03/10/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	03/10/2025	08:20
4	ICV	1	Water	NA	NA	20.2	7.01	03/10/2025	08:25
5	CCV1	1	Water	NA	NA	20.2	2.01	03/10/2025	08:30
6	Q1522-01	1	Water	NA	NA	20.8	5.91	03/10/2025	08:35
7	Q1528-01	1	Water	NA	NA	20.2	6.54	03/10/2025	08:44
8	Q1528-02	1	Water	NA	NA	20.4	8.40	03/10/2025	08:45
9	Q1528-03	1	Water	NA	NA	20.5	7.32	03/10/2025	08:50
10	Q1529-05	1	Water	NA	NA	20.3	6.14	03/10/2025	09:00
11	Q1529-06	1	Water	NA	NA	20.9	9.30	03/10/2025	09:10
12	Q1529-06DUP	1	Water	NA	NA	20.8	9.31	03/10/2025	09:11
13	CCV2	1	Water	NA	NA	20.2	12.02	03/10/2025	09:15

WORKLIST(Hardcopy Internal Chain)

VB134955

WorkList Name : ph w q1529

WorkList ID : 188130

Department : Wet-Chemistry

Date : 03-10-2025 08:03:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1522-01	TW-WTS-03	Water	pH	Cool 4 deg C	ENTA05	H31	03/06/2025	9040C
Q1528-01	C FERNOT	Water	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9040C
Q1528-02	A 22725	Water	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9040C
Q1528-03	C 22425	Water	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9040C
Q1529-05	A 60302	Water	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9040C
Q1529-06	A 60299	Water	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9040C

Date/Time 03/10/25 08:10

Raw Sample Received by: SA we

Raw Sample Relinquished by: SA we

Date/Time 03/10/25

Raw Sample Received by: SA we

Raw Sample Relinquished by: SA we