

Preparation Blank Summary

Client: ENTACT

SDG No.: Q1522

Project: 540 Degraw St, Brooklyn, NY - E9309

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	LB134945BL						
BOD5	mg/L	< 0.2000	0.2000	U	0.17	2.0	03/07/2025
Sample ID:	LB134959BLW						
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	03/07/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	03/07/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	03/07/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	03/07/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	03/07/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	03/07/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	03/07/2025
Sample ID:	LB134975BL						
TSS	mg/L	1	2.0000	J	1	4	03/10/2025