

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1522
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1513-05
Client ID:	DSN003DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	12.4		12.5		1	0.8		03/10/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1522
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1516-03
Client ID:	GAS-AUD-25-0026DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Flash Point	o F	+/-2	>212.0		>212.0		1	0		03/10/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: TW-WTS-04DUP	SDG No.: Q1522 Sample ID: Q1522-02 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	17.6		17.2		1	1.78		03/07/2025

Duplicate Sample Summary

Client: ENTACT	SDG No.: Q1522
Project: 540 Degraw St, Brooklyn, NY - E9309	Sample ID: Q1522-02
Client ID: TW-WTS-04MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	5.20		5.20		1	0		03/07/2025
Chloride	mg/L	+/-20	82.7	OR	82.7	OR	1	0		03/07/2025
Sulfate	mg/L	+/-20	375	OR	375	OR	1	0		03/07/2025
Nitrate	mg/L	+/-20	14.6	OR	14.7	OR	1	1		03/07/2025
Nitrite	mg/L	+/-20	11.2	OR	11.4	OR	1	2		03/07/2025
Fluoride	mg/L	+/-20	3.20		3.30		1	3		03/07/2025
Bromide	mg/L	+/-20	10.2		10.8		1	6		03/07/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1522
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1529-06
Client ID:	60299DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	9.30		9.31		1	0.11		03/10/2025