

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV44	Mercury	4.03	4.0	101	90 - 110	CV	03/17/2025	09:32	LB135051

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV15	Mercury	4.82	5.0	96	90 - 110	CV	03/17/2025	09:37	LB135051
CCV16	Mercury	4.83	5.0	97	90 - 110	CV	03/17/2025	10:02	LB135051
CCV17	Mercury	4.96	5.0	99	90 - 110	CV	03/17/2025	10:13	LB135051

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	506	510	99	95 - 105	P	03/12/2025	10:52	LB135011
	Copper	532	510	104	95 - 105	P	03/12/2025	10:52	LB135011
	Lead	1020	1000	102	95 - 105	P	03/12/2025	10:52	LB135011
	Nickel	517	530	98	95 - 105	P	03/12/2025	10:52	LB135011
	Zinc	1050	1000	105	95 - 105	P	03/12/2025	10:52	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.83	6.0	97	80 - 120	P	03/12/2025	10:56	LB135011
	Copper	22.9	20.0	114	80 - 120	P	03/12/2025	10:56	LB135011
	Lead	11.7	12.0	97	80 - 120	P	03/12/2025	10:56	LB135011
	Nickel	38.7	40.0	97	80 - 120	P	03/12/2025	10:56	LB135011
	Zinc	42.6	40.0	106	80 - 120	P	03/12/2025	10:56	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2470	2500	99	90 - 110	P	03/12/2025	11:29	LB135011
	Copper	1260	1250	101	90 - 110	P	03/12/2025	11:29	LB135011
	Lead	5060	5000	101	90 - 110	P	03/12/2025	11:29	LB135011
	Nickel	2450	2500	98	90 - 110	P	03/12/2025	11:29	LB135011
	Zinc	2530	2500	101	90 - 110	P	03/12/2025	11:29	LB135011
CCV02	Cadmium	2480	2500	99	90 - 110	P	03/12/2025	12:20	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	12:20	LB135011
	Lead	5010	5000	100	90 - 110	P	03/12/2025	12:20	LB135011
	Nickel	2460	2500	98	90 - 110	P	03/12/2025	12:20	LB135011
	Zinc	2600	2500	104	90 - 110	P	03/12/2025	12:20	LB135011
CCV03	Cadmium	2480	2500	99	90 - 110	P	03/12/2025	13:10	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	13:10	LB135011
	Lead	5000	5000	100	90 - 110	P	03/12/2025	13:10	LB135011
	Nickel	2480	2500	99	90 - 110	P	03/12/2025	13:10	LB135011
	Zinc	2600	2500	104	90 - 110	P	03/12/2025	13:10	LB135011
CCV04	Cadmium	2500	2500	100	90 - 110	P	03/12/2025	14:24	LB135011
	Copper	1290	1250	103	90 - 110	P	03/12/2025	14:24	LB135011
	Lead	5090	5000	102	90 - 110	P	03/12/2025	14:24	LB135011
	Nickel	2490	2500	100	90 - 110	P	03/12/2025	14:24	LB135011
	Zinc	2590	2500	103	90 - 110	P	03/12/2025	14:24	LB135011
CCV05	Cadmium	2460	2500	98	90 - 110	P	03/12/2025	15:25	LB135011
	Copper	1260	1250	101	90 - 110	P	03/12/2025	15:25	LB135011
	Lead	5010	5000	100	90 - 110	P	03/12/2025	15:25	LB135011
	Nickel	2440	2500	98	90 - 110	P	03/12/2025	15:25	LB135011
	Zinc	2530	2500	101	90 - 110	P	03/12/2025	15:25	LB135011
CCV06	Cadmium	2450	2500	98	90 - 110	P	03/12/2025	16:25	LB135011
	Copper	1260	1250	100	90 - 110	P	03/12/2025	16:25	LB135011
	Lead	4990	5000	100	90 - 110	P	03/12/2025	16:25	LB135011
	Nickel	2430	2500	97	90 - 110	P	03/12/2025	16:25	LB135011
	Zinc	2530	2500	101	90 - 110	P	03/12/2025	16:25	LB135011
CCV07	Cadmium	2460	2500	98	90 - 110	P	03/12/2025	17:25	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	17:25	LB135011
	Lead	5030	5000	101	90 - 110	P	03/12/2025	17:25	LB135011
	Nickel	2450	2500	98	90 - 110	P	03/12/2025	17:25	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q1522
Contract: ENTA05 Lab Code: CHEM Case No.: Q1522 SAS No.: Q1522
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2560	2500	102	90 - 110	P	03/12/2025	17:25	LB135011
CCV08	Cadmium	2490	2500	99	90 - 110	P	03/12/2025	18:23	LB135011
	Copper	1270	1250	102	90 - 110	P	03/12/2025	18:23	LB135011
	Lead	5060	5000	101	90 - 110	P	03/12/2025	18:23	LB135011
	Nickel	2470	2500	99	90 - 110	P	03/12/2025	18:23	LB135011
	Zinc	2580	2500	103	90 - 110	P	03/12/2025	18:23	LB135011
CCV09	Cadmium	2500	2500	100	90 - 110	P	03/12/2025	18:49	LB135011
	Copper	1280	1250	102	90 - 110	P	03/12/2025	18:49	LB135011
	Lead	5060	5000	101	90 - 110	P	03/12/2025	18:49	LB135011
	Nickel	2480	2500	99	90 - 110	P	03/12/2025	18:49	LB135011
	Zinc	2560	2500	102	90 - 110	P	03/12/2025	18:49	LB135011