



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.85	6.0	98	40 - 160	P	03/12/2025	11:05	LB135011
	Copper	22.6	20.0	113	40 - 160	P	03/12/2025	11:05	LB135011
	Lead	10.9	12.0	91	40 - 160	P	03/12/2025	11:05	LB135011
	Nickel	38.9	40.0	97	40 - 160	P	03/12/2025	11:05	LB135011
	Zinc	42.3	40.0	106	40 - 160	P	03/12/2025	11:05	LB135011
CRA	Mercury	0.23	0.2	114	40 - 160	CV	03/17/2025	09:41	LB135051