



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q1522				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q1522	SAS No.:	Q1522		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB44	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	09:34	LB135051	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB15	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	09:39	LB135051
CCB16	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	10:04	LB135051
CCB17	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	10:16	LB135051

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q1522

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1522

SAS No.: Q1522

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	11:01	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	11:01	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	11:01	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	11:01	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	11:01	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT			SDG No.:	Q1522				
Contract:	ENTA05	Lab Code:	CHEM	Case No.:	Q1522	SAS No.:	Q1522		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	11:34	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	11:34	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	11:34	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	11:34	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	11:34	LB135011
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	12:24	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	12:24	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	12:24	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	12:24	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	12:24	LB135011
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	13:14	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	13:14	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	13:14	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	13:14	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	13:14	LB135011
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	14:28	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	14:28	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	14:28	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	14:28	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	14:28	LB135011
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	15:29	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	15:29	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	15:29	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	15:29	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	15:29	LB135011
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	16:30	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	16:30	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	16:30	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	16:30	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	16:30	LB135011
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	17:30	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	17:30	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	17:30	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	17:30	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	17:30	LB135011
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	18:27	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	18:27	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	18:27	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	18:27	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1522</u>							
Contract: <u>ENTA05</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1522</u>		SAS No.: <u>Q1522</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	18:27	LB135011
CCB09	Cadmium	6.00	+/-6.00	U	6.00	P	03/12/2025	18:53	LB135011
	Copper	20.0	+/-20.0	U	20.0	P	03/12/2025	18:53	LB135011
	Lead	12.0	+/-12.0	U	12.0	P	03/12/2025	18:53	LB135011
	Nickel	40.0	+/-40.0	U	40.0	P	03/12/2025	18:53	LB135011
	Zinc	40.0	+/-40.0	U	40.0	P	03/12/2025	18:53	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1522</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1522</u>	SAS No.:	<u>Q1522</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number