

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-3.50	1.0	350	-5	7	03/12/2025	11:09	LB135011
	Copper	7.26	2.0	363	-18	22	03/12/2025	11:09	LB135011
	Lead	5.99			-12	12	03/12/2025	11:09	LB135011
	Nickel	1.62	2.0	81	-38	42	03/12/2025	11:09	LB135011
	Zinc	6.95			-40	40	03/12/2025	11:09	LB135011
ICSAB01	Cadmium	1000	972	103	826	1120	03/12/2025	11:14	LB135011
	Copper	509	511	100	434	588	03/12/2025	11:14	LB135011
	Lead	52.3	49.0	107	37	61	03/12/2025	11:14	LB135011
	Nickel	1000	954	105	810	1100	03/12/2025	11:14	LB135011
	Zinc	1080	952	113	809	1095	03/12/2025	11:14	LB135011