

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167038BS							
Cadmium	ug/L	100	92.4		92	80 - 120	P
Copper	ug/L	150	152		101	80 - 120	P
Lead	ug/L	500	475		95	80 - 120	P
Nickel	ug/L	250	235		94	80 - 120	P
Zinc	ug/L	100	105		105	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1522
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1522 **SAS No.:** Q1522

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167160BS Mercury	ug/L	4.0	4.77		119	80 - 120	CV