

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1522

Client: ENTACT

Analytical Method: 8270E DataFile: BF141942.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Low	Limits	
							Qual	Qual		High	RPD
PB167097BSD	Phenol	50	43.0	ug/L	86	6			20 (66)	160 (118)	20 (20)
	1,4-Dichlorobenzene	50	43.0	ug/L	86	6			70 (76)	130 (103)	20 (20)
	1,2,4-Trichlorobenzene	50	43.2	ug/L	86	4			70 (41)	130 (115)	20 (20)
	Naphthalene	50	42.1	ug/L	84	5			70 (64)	130 (107)	20 (20)