

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1522 SAS No.: Q1522 SDG NO.: Q1522

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/12/2025

Lab File ID: BF141931.D Time Analyzed: 10:08

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	605778	11.404	481594	14.039	422923	15.51
UPPER LIMIT	1211560	11.904	963188	14.539	845846	16.01
LOWER LIMIT	302889	10.904	240797	13.539	211462	15.01
EPA SAMPLE NO.						
01 TW-WTS-04	590417	11.40	355669	14.03	352171	15.51

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.