

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1522 SAS No.: Q1522 SDG No.: Q1522
 Instrument ID: MSVOA_N Calibration Date(s): 02/18/2025 02/18/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:09 14:18
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085772.D		RRF050 = VN085773.D		RRF010 = VN085775.D		
		RRF005 = VN085776.D		RRF001 = VN085777.D		RRF020 = VN085779.D		
COMPOUND	RRF100	RRF050	RRF010	RRF005	RRF001	RRF020	RRF	% RSD
Methyl tert-butyl Ether	1.747	1.658	1.714	1.560	1.438	1.659	1.629	7
Carbon Tetrachloride	0.573	0.528	0.583	0.545	0.533	0.538	0.550	4.1
Chloroform	1.120	1.070	1.269	1.164	1.171	1.141	1.156	5.7
1,1,1-Trichloroethane	1.012	0.967	1.089	1.058	1.030	1.012	1.028	4.1
Benzene	1.581	1.441	1.568	1.421	1.420	1.474	1.484	4.9
Toluene	0.993	0.902	0.924	0.820	0.689	0.891	0.870	12
Tetrachloroethene	0.384	0.361	0.420	0.391	0.393	0.375	0.387	5.2
Ethyl Benzene	2.105	1.904	1.830	1.665	1.424	1.838	1.794	12.8
m/p-Xylenes	0.809	0.741	0.725	0.610	0.481	0.722	0.682	17.2
o-Xylene	0.771	0.704	0.666	0.584	0.502	0.659	0.648	14.5
1,2-Dichloroethane-d4	0.698	0.613	0.588	0.767		0.575	0.648	12.6
Dibromofluoromethane	0.375	0.317	0.294	0.357		0.293	0.327	11.5
Toluene-d8	1.444	1.199	1.017	1.212		1.080	1.190	13.8
4-Bromofluorobenzene	0.503	0.419	0.323	0.370		0.346	0.392	18.2

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.