

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1522 SAS No.: Q1522 SDG No.: Q1522

Instrument ID: MSVOA_N Calibration Date/Time: 03/10/2025 12:21

Lab File ID: VN085914.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	1.629	1.871		14.86	20
Carbon Tetrachloride	0.550	0.609		10.73	20
Chloroform	1.156	1.197		3.55	20
1,1,1-Trichloroethane	1.028	1.086		5.64	20
Benzene	1.484	1.679		13.14	20
Toluene	0.870	1.062		22.07	20
Tetrachloroethene	0.387	0.411		6.2	20
Ethyl Benzene	1.794	2.144		19.51	20
m/p-Xylenes	0.682	0.863		26.54	20
o-Xylene	0.648	0.796		22.84	20
1,2-Dichloroethane-d4	0.648	0.628		-3.09	20
Dibromofluoromethane	0.327	0.328		0.31	20
Toluene-d8	1.190	1.216		2.18	20
4-Bromofluorobenzene	0.392	0.455		16.07	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.