

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN021825

Review By	John Carlone	Review On	2/19/2025 9:44:55 AM
Supervise By	Maresh Dadoda	Supervise On	2/19/2025 1:34:51 PM
SubDirectory	VN021825	HP Acquire Method	HP Processing Method 82N021825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133068 VP133070,VP133071,VP133072,VP133073,VP133074,VP133075 VP133076

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085771.D	18 Feb 2025 10:35		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN085772.D	18 Feb 2025 11:09	Comp.#43 is on Linear Regression	JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN085773.D	18 Feb 2025 11:32	Comp.#56,58,70,86 is on Quadratic Regression	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN085774.D	18 Feb 2025 11:56	Not used	JC\MD	Not Ok
5	VSTDICC010	VSTDICC010	VN085775.D	18 Feb 2025 12:20		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN085776.D	18 Feb 2025 12:43	%D failed for Comp. #58 in 01PPB, 05PPB and 20PPB	JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN085777.D	18 Feb 2025 13:07		JC\MD	Ok,M
8	IBLK	IBLK	VN085778.D	18 Feb 2025 13:54		JC\MD	Ok
9	VSTDICC020	VSTDICC020	VN085779.D	18 Feb 2025 14:18		JC\MD	Ok,M
10	VSTDICV050	ICVVN021825	VN085780.D	18 Feb 2025 16:28	ICV Failed for comp. #95	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN031025

Review By	John Carlone	Review On	3/11/2025 9:56:40 AM
Supervise By	Maresh Dadoda	Supervise On	3/11/2025 3:08:56 PM
SubDirectory	VN031025	HP Acquire Method	HP Processing Method 82N021825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133238		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133239,VP133240		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085913.D	10 Mar 2025 10:54		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085914.D	10 Mar 2025 12:21	CCAL failed high for comp. #68,69	JC\MD	Ok,M
3	VN0310MBL01	VN0310MBL01	VN085915.D	10 Mar 2025 12:58		JC\MD	Ok
4	VN0310WBL01	VN0310WBL01	VN085916.D	10 Mar 2025 13:22		JC\MD	Ok
5	VN0310WBS01	VN0310WBS01	VN085917.D	10 Mar 2025 16:12	BS failed high for comp. #68	JC\MD	Ok,M
6	VN0310WBSD01	VN0310WBSD01	VN085918.D	10 Mar 2025 16:47		JC\MD	Ok,M
7	Q1514-07	ENV-105-GW01	VN085919.D	10 Mar 2025 17:11	vial A pH<2	JC\MD	Ok
8	Q1514-08	ENV-103-GW01	VN085920.D	10 Mar 2025 17:35	vial A pH<2	JC\MD	Ok
9	Q1514-09	FB03062025	VN085921.D	10 Mar 2025 17:59	vial A pH<2 FB;Hit of comp.#16	JC\MD	ReRun
10	Q1514-10	TB03062025	VN085922.D	10 Mar 2025 18:23	vial A pH<2 TB	JC\MD	Ok
11	Q1516-03	GAS-AUD-25-0026	VN085923.D	10 Mar 2025 18:47	vial A pH<2	JC\MD	Ok
12	Q1525-01	MW10	VN085924.D	10 Mar 2025 19:12	vial A pH<2 RT shifted; CCAL failed high for comp. #68,69;	JC\MD	Not Ok
13	Q1522-02	TW-WTS-04	VN085925.D	10 Mar 2025 19:35	vial A pH<2	JC\MD	Ok
14	VSTDCCC050	VSTDCCC050EC	VN085926.D	10 Mar 2025 19:59		JC\MD	Ok,M

M : Manual Integration