

Surrogate Summary

SDG No.: Q1522

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1522-02	TW-WTS-04	1,2-Dichloroethane-d4	50	55.3	111	70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100	70 (75)	130 (124)
		Toluene-d8	50	47.9	96	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.7	97	70 (77)	130 (121)
VN0310WBL01	VN0310WBL01	1,2-Dichloroethane-d4	50	54.6	109	70 (74)	130 (125)
		Dibromofluoromethane	50	50.6	101	70 (75)	130 (124)
		Toluene-d8	50	47.0	94	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.0	90	70 (77)	130 (121)
VN0310WBS01	VN0310WBS01	1,2-Dichloroethane-d4	50	50.3	101	70 (74)	130 (125)
		Dibromofluoromethane	50	49.3	99	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.4	109	70 (77)	130 (121)
VN0310WBSD0	VN0310WBSD01	1,2-Dichloroethane-d4	50	48.9	98	70 (74)	130 (125)
		Dibromofluoromethane	50	47.2	94	70 (75)	130 (124)
		Toluene-d8	50	46.9	94	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)

