

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1522 SAS No.: Q1522 SDG NO.: Q1522
 Lab File ID: VN085914.D Date Analyzed: 03/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:21
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	199203	8.22	319135	9.10	288321	11.87
UPPER LIMIT	398406	8.724	638270	9.6	576642	12.365
LOWER LIMIT	99601.5	7.724	159568	8.6	144161	11.365
EPA SAMPLE NO.						
TW-WTS-04	197312	8.22	371245	9.10	322337	11.87
VN0310WBL01	185188	8.22	337149	9.10	297684	11.87
VN0310WBS01	200845	8.22	334623	9.10	292041	11.87
VN0310WBSD01	183155	8.22	304090	9.10	267080	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.