



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1523
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	WC-A1-01-G							
Q1523-01	WC-A1-01-G	TCLP	Benzene	12.3		0.15	5.00	ug/L
			Total Voc :	12.3				
			Total Concentration:	12.3				
Client ID:	WC-A1-02-G							
Q1523-04	WC-A1-02-G	TCLP	Benzene	59.7		0.15	5.00	ug/L
			Total Voc :	59.7				
			Total Concentration:	59.7				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-01-G	SDG No.:	Q1523
Lab Sample ID:	Q1523-01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085950.D	1		03/11/25 22:41	VN031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	5.00	ug/L
71-43-2	Benzene	12.3		0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	5.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	5.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	5.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	58.8		70 (74) - 130 (125)	118%	SPK: 50
1868-53-7	Dibromofluoromethane	41.1		70 (75) - 130 (124)	82%	SPK: 50
2037-26-5	Toluene-d8	48.5		70 (86) - 130 (113)	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.9		70 (77) - 130 (121)	110%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	159000	8.224			
540-36-3	1,4-Difluorobenzene	305000	9.1			
3114-55-4	Chlorobenzene-d5	277000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	133000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-02-G	SDG No.:	Q1523
Lab Sample ID:	Q1523-04	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085951.D	1		03/11/25 23:05	VN031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	5.00	ug/L
71-43-2	Benzene	59.7		0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	5.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	5.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	5.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.7		70 (74) - 130 (125)	111%	SPK: 50
1868-53-7	Dibromofluoromethane	38.3		70 (75) - 130 (124)	77%	SPK: 50
2037-26-5	Toluene-d8	47.6		70 (86) - 130 (113)	95%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.9		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	183000	8.224			
540-36-3	1,4-Difluorobenzene	343000	9.1			
3114-55-4	Chlorobenzene-d5	297000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	139000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q1523	OrderDate:	3/7/2025 10:35:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	I31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1523-01	WC-A1-01-G	TCLP	TCLP VOA	8260D	03/06/25		03/11/25	03/07/25
Q1523-04	WC-A1-02-G	TCLP	TCLP VOA	8260D	03/06/25		03/11/25	03/07/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

A

B

C

D

Hit Summary Sheet SW-846

SDG No.: Q1523
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:	03/11/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/11/25	
Client Sample ID:	PB167049TB		SDG No.:	Q1523	
Lab Sample ID:	PB167049TB		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141934.D	1	03/11/25 11:15	03/12/25 11:36	PB167089

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/Kg
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/Kg
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/Kg
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/Kg
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/Kg
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/Kg
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/Kg
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/Kg
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/Kg
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/Kg
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/Kg
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/Kg
SURROGATES						
367-12-4	2-Fluorophenol	142		15 (10) - 110 (139)	95%	SPK: 150
13127-88-3	Phenol-d6	135		15 (10) - 110 (134)	90%	SPK: 150
4165-60-0	Nitrobenzene-d5	97.7		30 (49) - 130 (133)	98%	SPK: 100
321-60-8	2-Fluorobiphenyl	98.3		30 (52) - 130 (132)	98%	SPK: 100
118-79-6	2,4,6-Tribromophenol	144		15 (44) - 110 (137)	96%	SPK: 150
1718-51-0	Terphenyl-d14	90.8		30 (48) - 130 (125)	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	142000	6.881			
1146-65-2	Naphthalene-d8	564000	8.157			
15067-26-2	Acenaphthene-d10	314000	9.916			
1517-22-2	Phenanthrene-d10	545000	11.398			
1719-03-5	Chrysene-d12	397000	14.033			
1520-96-3	Perylene-d12	371000	15.51			

Report of Analysis

Client:	ENTACT		Date Collected:	03/11/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/11/25	
Client Sample ID:	PB167049TB		SDG No.:	Q1523	
Lab Sample ID:	PB167049TB		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141934.D	1	03/11/25 11:15	03/12/25 11:36	PB167089

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-01-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-03		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141923.D	1	03/11/25 11:15	03/11/25 18:42	PB167089

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	124		15 (10) - 110 (139)	83%	SPK: 150
13127-88-3	Phenol-d6	109		15 (10) - 110 (134)	72%	SPK: 150
4165-60-0	Nitrobenzene-d5	95.3		30 (49) - 130 (133)	95%	SPK: 100
321-60-8	2-Fluorobiphenyl	103		30 (52) - 130 (132)	103%	SPK: 100
118-79-6	2,4,6-Tribromophenol	135		15 (44) - 110 (137)	90%	SPK: 150
1718-51-0	Terphenyl-d14	82.4		30 (48) - 130 (125)	82%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	133000	6.881			
1146-65-2	Naphthalene-d8	499000	8.163			
15067-26-2	Acenaphthene-d10	254000	9.91			
1517-22-2	Phenanthrene-d10	415000	11.398			
1719-03-5	Chrysene-d12	364000	14.033			
1520-96-3	Perylene-d12	276000	15.504			

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-01-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-03		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141923.D	1	03/11/25 11:15	03/11/25 18:42	PB167089

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-02-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-06		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG064100.D	1	03/11/25 11:15	03/11/25 17:59	PB167089

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	130		15 (10) - 110 (139)	87%	SPK: 150
13127-88-3	Phenol-d6	117		15 (10) - 110 (134)	78%	SPK: 150
4165-60-0	Nitrobenzene-d5	110		30 (49) - 130 (133)	110%	SPK: 100
321-60-8	2-Fluorobiphenyl	94.2		30 (52) - 130 (132)	94%	SPK: 100
118-79-6	2,4,6-Tribromophenol	159		15 (44) - 110 (137)	106%	SPK: 150
1718-51-0	Terphenyl-d14	97.9		30 (48) - 130 (125)	98%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	37300	7.865			
1146-65-2	Naphthalene-d8	156000	10.65			
15067-26-2	Acenaphthene-d10	104000	14.486			
1517-22-2	Phenanthrene-d10	217000	17.224			
1719-03-5	Chrysene-d12	226000	21.455			
1520-96-3	Perylene-d12	245000	24.469			

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-02-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-06		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG064100.D	1	03/11/25 11:15	03/11/25 17:59	PB167089

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q1523	OrderDate:	3/7/2025 10:35:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	I31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1523-03	WC-A1-01-C	TCLP	TCLP BNA	8270E	03/06/25	03/11/25	03/11/25	03/07/25
Q1523-06	WC-A1-02-C	TCLP	TCLP BNA	8270E	03/06/25	03/11/25	03/11/25	03/07/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1523

Order ID: Q1523

Client: ENTACT

Project ID: 540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A

B

C

D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:		
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/11/25	
Client Sample ID:	PB167049TB		SDG No.:	Q1523	
Lab Sample ID:	PB167049TB		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094685.D	1	03/11/25 11:54	03/13/25 17:47	PB167091

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	23.5		30 (43) - 150 (140)	117%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.1		30 (77) - 150 (126)	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-01-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-03		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094661.D	1	03/11/25 11:54	03/12/25 21:41	PB167091

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	25.1		30 (43) - 150 (140)	126%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.0		30 (77) - 150 (126)	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-02-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-06		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094662.D	1	03/11/25 11:54	03/12/25 21:55	PB167091

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	26.9		30 (43) - 150 (140)	134%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.4		30 (77) - 150 (126)	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1523	OrderDate:	3/7/2025 10:35:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	I31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1523-02	WC-A1-01-C	SOIL	PCB	8082A	03/06/25	03/10/25	03/10/25	03/07/25
Q1523-03	WC-A1-01-C	TCLP	TCLP Herbicide TCLP Pesticide	8151A 8081B	03/06/25	03/11/25 03/11/25	03/12/25 03/12/25	03/07/25
Q1523-05	WC-A1-02-C	SOIL	PCB	8082A	03/06/25	03/10/25	03/10/25	03/07/25
Q1523-06	WC-A1-02-C	TCLP	TCLP Herbicide TCLP Pesticide	8151A 8081B	03/06/25	03/11/25 03/11/25	03/12/25 03/12/25	03/07/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1523

Order ID: Q1523

Client: ENTACT

Project ID: 540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								

Total Concentration: 0.000



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-01-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-02		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	78.6	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070397.D	1	03/10/25 08:35	03/10/25 14:01	PB167043

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.30	U	4.30	21.6	ug/kg
11104-28-2	Aroclor-1221	8.10	U	8.10	21.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	21.6	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30	21.6	ug/kg
12672-29-6	Aroclor-1248	10.0	U	10.0	21.6	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	21.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	21.6	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	21.6	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	21.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.7		30 (32) - 150 (144)	114%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.8		30 (32) - 150 (175)	79%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-02-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-05		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	78	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070400.D	1	03/10/25 08:35	03/10/25 14:49	PB167043

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.30	U	4.30	21.8	ug/kg
11104-28-2	Aroclor-1221	8.20	U	8.20	21.8	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	21.8	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30	21.8	ug/kg
12672-29-6	Aroclor-1248	10.1	U	10.1	21.8	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	21.8	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	21.8	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	21.8	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	21.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.0		30 (32) - 150 (144)	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.4		30 (32) - 150 (175)	77%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1523	OrderDate:	3/7/2025 10:35:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	I31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1523-02	WC-A1-01-C	SOIL			03/06/25			03/07/25
			PCB	8082A		03/10/25	03/10/25	
Q1523-03	WC-A1-01-C	TCLP			03/06/25			03/07/25
			TCLP Herbicide	8151A		03/11/25	03/12/25	
			TCLP Pesticide	8081B		03/11/25	03/12/25	
Q1523-05	WC-A1-02-C	SOIL			03/06/25			03/07/25
			PCB	8082A		03/10/25	03/10/25	
Q1523-06	WC-A1-02-C	TCLP			03/06/25			03/07/25
			TCLP Herbicide	8151A		03/11/25	03/12/25	
			TCLP Pesticide	8081B		03/11/25	03/12/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1523

Order ID: Q1523

Client: ENTACT

Project ID: 540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A
B
C
D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:		
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/11/25	
Client Sample ID:	PB167049TB		SDG No.:	Q1523	
Lab Sample ID:	PB167049TB		Matrix:	TCLP	
Analytical Method:	SW8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS029400.D	1	03/11/25 11:30	03/12/25 11:09	PB167090

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	9.20	U	9.20	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	7.80	U	7.80	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	538		70 (39) - 130 (175)	108%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-01-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-03		Matrix:	TCLP	
Analytical Method:	SW8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS029414.D	1	03/11/25 11:30	03/12/25 17:03	PB167090

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	9.20	U	9.20	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	7.80	U	7.80	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	490		70 (39) - 130 (175)	98%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	03/06/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	03/07/25	
Client Sample ID:	WC-A1-02-C		SDG No.:	Q1523	
Lab Sample ID:	Q1523-06		Matrix:	TCLP	
Analytical Method:	SW8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS029415.D	1	03/11/25 11:30	03/12/25 17:27	PB167090

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	9.20	U	9.20	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	7.80	U	7.80	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	511		70 (39) - 130 (175)	102%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1523	OrderDate:	3/7/2025 10:35:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	I31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1523-02	WC-A1-01-C	SOIL			03/06/25			03/07/25
			PCB	8082A		03/10/25	03/10/25	
Q1523-03	WC-A1-01-C	TCLP			03/06/25			03/07/25
			TCLP Herbicide	8151A		03/11/25	03/12/25	
			TCLP Pesticide	8081B		03/11/25	03/12/25	
Q1523-05	WC-A1-02-C	SOIL			03/06/25			03/07/25
			PCB	8082A		03/10/25	03/10/25	
Q1523-06	WC-A1-02-C	TCLP			03/06/25			03/07/25
			TCLP Herbicide	8151A		03/11/25	03/12/25	
			TCLP Pesticide	8081B		03/11/25	03/12/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.:	Q1523	Order ID:	Q1523
Client:	ENTACT	Project ID:	540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WC-A1-01-C								
Q1523-03	WC-A1-01-C	TCLP	Barium	437	J	62.8	500	ug/L
Q1523-03	WC-A1-01-C	TCLP	Chromium	16.7	J	6.60	50.0	ug/L
Client ID : WC-A1-02-C								
Q1523-06	WC-A1-02-C	TCLP	Barium	382	J	62.8	500	ug/L
Q1523-06	WC-A1-02-C	TCLP	Chromium	20.5	J	6.60	50.0	ug/L



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-01-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-03	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	34.8	U	1	34.8	100	ug/L	03/11/25 12:05	03/11/25 17:16	SW6010	SW3050
7440-39-3	Barium	437	J	1	62.8	500	ug/L	03/11/25 12:05	03/11/25 17:16	SW6010	SW3050
7440-43-9	Cadmium	0.94	U	1	0.94	30.0	ug/L	03/11/25 12:05	03/11/25 17:16	SW6010	SW3050
7440-47-3	Chromium	16.7	J	1	6.60	50.0	ug/L	03/11/25 12:05	03/11/25 17:16	SW6010	SW3050
7439-92-1	Lead	35.1	U	1	35.1	60.0	ug/L	03/11/25 12:05	03/11/25 17:16	SW6010	SW3050
7439-97-6	Mercury	0.81	UN	1	0.81	2.00	ug/L	03/11/25 15:10	03/12/25 11:32	SW7470A	
7782-49-2	Selenium	58.8	U	1	58.8	100	ug/L	03/11/25 12:05	03/11/25 17:16	SW6010	SW3050
7440-22-4	Silver	5.80	U	1	5.80	50.0	ug/L	03/11/25 12:05	03/11/25 17:16	SW6010	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP-FULL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-02-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-06	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	34.8	U	1	34.8	100	ug/L	03/11/25 12:05	03/11/25 17:20	SW6010	SW3050
7440-39-3	Barium	382	J	1	62.8	500	ug/L	03/11/25 12:05	03/11/25 17:20	SW6010	SW3050
7440-43-9	Cadmium	0.94	U	1	0.94	30.0	ug/L	03/11/25 12:05	03/11/25 17:20	SW6010	SW3050
7440-47-3	Chromium	20.5	J	1	6.60	50.0	ug/L	03/11/25 12:05	03/11/25 17:20	SW6010	SW3050
7439-92-1	Lead	35.1	U	1	35.1	60.0	ug/L	03/11/25 12:05	03/11/25 17:20	SW6010	SW3050
7439-97-6	Mercury	0.81	UN	1	0.81	2.00	ug/L	03/11/25 15:10	03/12/25 11:34	SW7470A	
7782-49-2	Selenium	58.8	U	1	58.8	100	ug/L	03/11/25 12:05	03/11/25 17:20	SW6010	SW3050
7440-22-4	Silver	5.80	U	1	5.80	50.0	ug/L	03/11/25 12:05	03/11/25 17:20	SW6010	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP-FULL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q1523	OrderDate:	3/7/2025 10:35:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	I31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1523-03	WC-A1-01-C	TCLP			03/06/25			03/07/25
			TCLP ICP Metals	6010D		03/11/25	03/11/25	
			TCLP Mercury	7470A		03/11/25	03/12/25	
Q1523-06	WC-A1-02-C	TCLP			03/06/25			03/07/25
			TCLP ICP Metals	6010D		03/11/25	03/11/25	
			TCLP Mercury	7470A		03/11/25	03/12/25	



SAMPLE DATA

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25 10:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-01-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-02	Matrix:	SOIL
		% Solid:	78.6

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Oil and Grease	3030		1	4.13	31.7	mg/Kg		03/11/25 09:45	SW9071B
Paint Filter	1.00	U	1	1.00	1.00	ml/100gm		03/10/25 11:12	9095B
pH	12.3	H	1	0	0	pH		03/10/25 09:55	9045D
TS	78.9		1	1.00	5.00	%		03/07/25 12:00	SM 2540 B-15
TVS	2.60	J	1	1.00	10.0	%		03/07/25 15:30	160.4

Comments: pH result reported at temperature 23.9 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25 10:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-01-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-03	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	12.3	H	1	0	0	pH		03/10/25 09:55	9045D
Ignitability	NO		1	0	0	oC		03/10/25 09:52	1030
Reactive Cyanide	0.063	J	1	0.043	0.25	mg/Kg	03/10/25 11:00	03/10/25 15:44	9012B
Reactive Sulfide	4.79	J	1	0.19	10.0	mg/Kg	03/10/25 13:30	03/11/25 08:53	9034

Comments: pH result reported at temperature 23.9 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25 11:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-02-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-05	Matrix:	SOIL
		% Solid:	78

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Oil and Grease	3440		1	4.16	32.0	mg/Kg		03/11/25 09:45	SW9071B
Paint Filter	1.00	U	1	1.00	1.00	ml/100gm		03/10/25 11:20	9095B
pH	12.3	H	1	0	0	pH		03/10/25 10:10	9045D
TS	78.3		1	1.00	5.00	%		03/07/25 12:00	SM 2540 B-15
TVS	2.90	J	1	1.00	10.0	%		03/07/25 15:30	160.4

Comments: pH result reported at temperature 24.0 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25 11:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-02-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-06	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	12.3	H	1	0	0	pH		03/10/25 10:10	9045D
Ignitability	NO		1	0	0	oC		03/10/25 10:00	1030
Reactive Cyanide	0.071	J	1	0.043	0.24	mg/Kg	03/10/25 11:00	03/10/25 15:44	9012B
Reactive Sulfide	6.35	J	1	0.19	10.0	mg/Kg	03/10/25 13:30	03/11/25 08:56	9034

Comments: pH result reported at temperature 24.0 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25 10:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-01-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-07	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
ASTM Ammonia	0.47		1	0.045	0.10	mg/L	03/12/25 08:45	03/12/25 17:20	SM 4500-NH3 B plus NH3 G-11
ASTM COD	23.6		1	2.35	10.0	mg/L		03/12/25 15:02	SM 5220 D-11
ASTM Oil and Grease	0.40	U	1	0.40	5.00	mg/L		03/12/25 10:30	SW1664A
ASTM TS	1200		1	1.00	5.00	mg/L		03/11/25 10:00	SM 2540 B-15

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	03/06/25 11:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	03/07/25
Client Sample ID:	WC-A1-02-C	SDG No.:	Q1523
Lab Sample ID:	Q1523-08	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
ASTM Ammonia	0.70		1	0.045	0.10	mg/L	03/12/25 08:45	03/12/25 17:20	SM 4500-NH3 B plus NH3 G-11
ASTM COD	19.6		1	2.35	10.0	mg/L		03/12/25 15:02	SM 5220 D-11
ASTM Oil and Grease	0.40	J	1	0.40	5.00	mg/L		03/12/25 10:30	SW1664A
ASTM TS	1210		1	1.00	5.00	mg/L		03/11/25 10:00	SM 2540 B-15

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q1523	OrderDate:	3/7/2025 10:35:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	I31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1523-02	WC-A1-01-C	SOIL			03/06/25 10:00			03/07/25
			Oil and Grease	9071B			03/11/25 09:45	
			Paint Filter	9095B			03/10/25 11:12	
			pH	9045D			03/10/25 09:55	
			TS	SM2540 B			03/07/25 12:00	
			TVS	160.4			03/07/25 15:30	
Q1523-03	WC-A1-01-C	SOIL			03/06/25 10:00			03/07/25
			Corrosivity	9045D			03/10/25 09:55	
			Ignitability	1030			03/10/25 09:52	
			Reactive Cyanide	9012B		03/10/25	03/10/25 15:44	
			Reactive Sulfide	9034		03/10/25	03/11/25 08:53	
Q1523-05	WC-A1-02-C	SOIL			03/06/25 11:00			03/07/25
			Oil and Grease	9071B			03/11/25 09:45	
			Paint Filter	9095B			03/10/25 11:20	
			pH	9045D			03/10/25 10:10	

LAB CHRONICLE

Q1523-06	WC-A1-02-C	SOIL	TS	SM2540 B		03/07/25 12:00
			TVS	160.4		03/07/25 15:30
					03/06/25 11:00	03/07/25
			Corrosivity	9045D		03/10/25 10:10
			Ignitability	1030		03/10/25 10:00
Q1523-07	WC-A1-01-C	WATER	Reactive Cyanide	9012B	03/10/25	03/10/25 15:44
			Reactive Sulfide	9034	03/10/25	03/11/25 08:56
					03/06/25 10:00	03/07/25
			ASTM Ammonia	SM4500-NH3	03/12/25	03/12/25 17:20
			ASTM COD	SM5220 D		03/12/25 15:02
Q1523-08	WC-A1-02-C	WATER	ASTM Oil and Grease	1664A		03/12/25 10:30
			ASTM TS	SM2540 B		03/11/25 10:00
					03/06/25 11:00	03/07/25
			ASTM Ammonia	SM4500-NH3	03/12/25	03/12/25 17:20
			ASTM COD	SM5220 D		03/12/25 15:02
			ASTM Oil and Grease	1664A		03/12/25 10:30
			ASTM TS	SM2540 B		03/11/25 10:00