



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134956

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	03/10/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	03/10/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	03/10/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134957

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	03/10/2025
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	03/10/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	03/10/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134969

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.097	0.099	98	85-115	03/10/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	03/10/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	03/10/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	03/10/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	03/10/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135006

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 ASTM COD	mg/L	49.329	50	99	95-105	03/12/2025
Sample ID: CCV2 ASTM COD	mg/L	48.339	50	97	95-105	03/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135013

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	0.99	1	99	90-110	03/12/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	0.96	1	96	90-110	03/12/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	0.98	1	98	90-110	03/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135013

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------