

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB134976

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134976BS							
Oil and Grease	mg/Kg	100	89.9		90	1	80-120	03/11/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB134993

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134993BS							
ASTM Oil and Grease	mg/L	20.0	16.8		84	1	78-114	03/12/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135006

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135006BS							
ASTM COD	mg/L	50	48.3		97	1	90-110	03/12/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1523

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135013

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167083BS							
ASTM Ammonia	mg/L	1	0.95		95	1	90-110	03/12/2025