



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1523
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1523 **SAS No.:** Q1523
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.0	20.0	95	40 - 160	P	03/11/2025	11:02	LB134987
	Barium	95.1	100	95	40 - 160	P	03/11/2025	11:02	LB134987
	Cadmium	5.81	6.0	97	40 - 160	P	03/11/2025	11:02	LB134987
	Chromium	9.83	10.0	98	40 - 160	P	03/11/2025	11:02	LB134987
	Lead	11.9	12.0	99	40 - 160	P	03/11/2025	11:02	LB134987
	Selenium	21.9	20.0	110	40 - 160	P	03/11/2025	11:02	LB134987
	Silver	10.3	10.0	103	40 - 160	P	03/11/2025	11:02	LB134987
CRA	Mercury	0.22	0.2	109	40 - 160	CV	03/12/2025	11:15	LB134996