



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q1523			
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q1523	SAS No.:	Q1523	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB38	Mercury	0.20	+/-0.20	U	0.20	CV	03/12/2025	11:02	LB134996

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1523</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1523</u>	SAS No.: <u>Q1523</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB86	Mercury	0.20	+/-0.20	U	0.20	CV	03/12/2025	11:12	LB134996
CCB87	Mercury	0.20	+/-0.20	U	0.20	CV	03/12/2025	11:46	LB134996
CCB88	Mercury	0.20	+/-0.20	U	0.20	CV	03/12/2025	12:56	LB134996
CCB89	Mercury	0.20	+/-0.20	U	0.20	CV	03/12/2025	13:20	LB134996
CCB90	Mercury	0.20	+/-0.20	U	0.20	CV	03/12/2025	13:48	LB134996

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q1523			
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q1523	SAS No.:	Q1523	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	10:58	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	10:58	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	10:58	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	10:58	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	10:58	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	10:58	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	10:58	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1523</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1523</u>	SAS No.: <u>Q1523</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	11:34	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	11:34	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	11:34	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	11:34	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	11:34	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	11:34	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	11:34	LB134987
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	12:23	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	12:23	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	12:23	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	12:23	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	12:23	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	12:23	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	12:23	LB134987
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	13:14	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	13:14	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	13:14	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	13:14	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	13:14	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	13:14	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	13:14	LB134987
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	14:13	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	14:13	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	14:13	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	14:13	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	14:13	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	14:13	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	14:13	LB134987
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	15:04	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	15:04	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	15:04	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	15:04	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	15:04	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	15:04	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	15:04	LB134987
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	15:54	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	15:54	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	15:54	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	15:54	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1523							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1523	SAS No.: Q1523						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	15:54	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	15:54	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	15:54	LB134987
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	16:46	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	16:46	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	16:46	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	16:46	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	16:46	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	16:46	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	16:46	LB134987
	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	17:38	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	17:38	LB134987
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	17:38	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	17:38	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	17:38	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	17:38	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	17:38	LB134987
	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	18:30	LB134987
CCB09	Barium	100	+/-100	U	100	P	03/11/2025	18:30	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	18:30	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	18:30	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	18:30	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	18:30	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	18:30	LB134987
CCB10	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	19:18	LB134987
	Barium	100	+/-100	U	100	P	03/11/2025	19:18	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	19:18	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	19:18	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	19:18	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	19:18	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	19:18	LB134987
	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	19:31	LB134987
CCB11	Barium	100	+/-100	U	100	P	03/11/2025	19:31	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	19:31	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	19:31	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	19:31	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	19:31	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	19:31	LB134987
CCB12	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	19:49	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1523</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1523</u>	SAS No.:	<u>Q1523</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Barium	100	+/-100	U	100	P	03/11/2025	19:49	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	19:49	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	19:49	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	19:49	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	19:49	LB134987
	Silver	10.0	+/-10.0	U	10.0	P	03/11/2025	19:49	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1523</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1523</u>	SAS No.:	<u>Q1523</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number