

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q1523
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1523 **SAS No.:** Q1523
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	6.46			-20	20	03/11/2025	11:06	LB134987
	Barium	5.32	6.0	89	-94	106	03/11/2025	11:06	LB134987
	Cadmium	-2.31	1.0	231	-5	7	03/11/2025	11:06	LB134987
	Chromium	58.2	52.0	112	42	62	03/11/2025	11:06	LB134987
	Lead	3.81			-12	12	03/11/2025	11:06	LB134987
	Selenium	-16.3			-20	20	03/11/2025	11:06	LB134987
	Silver	-0.94			-10	10	03/11/2025	11:06	LB134987
ICSAB01	Arsenic	114	104	110	88.4	120	03/11/2025	11:11	LB134987
	Barium	487	537	91	437	637	03/11/2025	11:11	LB134987
	Cadmium	1020	972	105	826	1120	03/11/2025	11:11	LB134987
	Chromium	567	542	105	460	624	03/11/2025	11:11	LB134987
	Lead	55.7	49.0	114	37	61	03/11/2025	11:11	LB134987
	Selenium	32.4	46.0	70	26	66	03/11/2025	11:11	LB134987
	Silver	220	201	110	170	232	03/11/2025	11:11	LB134987