

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: ENTACT **SDG No.:** Q1523
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1523 **SAS No.:** Q1523
Matrix: Water **Level:** LOW **Client ID:** COMP-3A
Sample ID: Q1537-07 **Spiked ID:** Q1537-07A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	38.7		2.00	U	40.0	97		CV