

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1523
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1523 **SAS No.:** Q1523

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167092BS							
Arsenic	ug/L	4000	3800		95	80 - 120	P
Barium	ug/L	1000	866		87	80 - 120	P
Cadmium	ug/L	1000	916		92	80 - 120	P
Chromium	ug/L	2000	1890		94	80 - 120	P
Lead	ug/L	5000	4610		92	80 - 120	P
Selenium	ug/L	10000	9220		92	80 - 120	P
Silver	ug/L	380	362		95	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q1523

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1523

SAS No.: Q1523

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167093BS Mercury	ug/L	4.0	4.25		106	80 - 120	CV