

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167043BS

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

Lab Sample ID: PB167043BS Date(s) Analyzed: 03/10/2025 03/10/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP070396.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.677	5.627	5.727	159	157	6.17
COLUMN 1	2	5.699	5.649	5.749	158		
	3	5.761	5.711	5.811	157		
	4	5.859	5.809	5.909	162		
	5	6.152	6.102	6.202	148		
	1	4.913	4.863	4.963	166		
COLUMN 2	2	4.932	4.882	4.982	166	167	
	3	5.11	5.06	5.16	170		
	4	5.152	5.102	5.202	166		
	5	5.366	5.316	5.416	167		
Aroclor-1260	1	7.27	7.22	7.32	164	151	2.61
COLUMN 1	2	7.524	7.474	7.574	157		
	3	7.883	7.833	7.933	142		
	4	8.107	8.057	8.157	148		
	5	8.428	8.378	8.478	146		
	1	6.404	6.354	6.454	158		
COLUMN 2	2	6.592	6.542	6.642	160	155	
	3	6.745	6.695	6.795	150		
	4	7.218	7.168	7.268	151		
	5	7.459	7.409	7.509	155		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

WC-A1-01-CMS

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1523

SAS No.: Q1523

SDG NO.: Q1523

Lab Sample ID: Q1523-02MS

Date(s) Analyzed: 03/10/2025

03/10/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP070398.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.677	5.627	5.727	98.7	138	19.61
COLUMN 1	2	5.701	5.651	5.751	142		
	3	5.76	5.71	5.81	183		
	4	5.858	5.808	5.908	154		
	5	6.149	6.099	6.199	110		
	1	4.915	4.865	4.965	116		
COLUMN 2	2	4.934	4.884	4.984	141	168	
	3	5.112	5.062	5.162	147		
	4	5.153	5.103	5.203	145		
	5	5.381	5.331	5.431	291		
Aroclor-1260	1	7.271	7.221	7.321	151	167	14.79
COLUMN 1	2	7.525	7.475	7.575	207		
	3	7.884	7.834	7.934	145		
	4	8.108	8.058	8.158	162		
	5	8.429	8.379	8.479	171		
	1	6.406	6.356	6.456	145		
COLUMN 2	2	6.594	6.544	6.644	142	144	
	3	6.747	6.697	6.797	133		
	4	7.22	7.17	7.27	156		
	5	7.461	7.411	7.511	142		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

WC-A1-01-CMSD

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1523

SAS No.: Q1523

SDG NO.: Q1523

Lab Sample ID: Q1523-02MSD

Date(s) Analyzed: 03/10/2025

03/10/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP070399.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.674	5.624	5.724	95.1	138	14.77
COLUMN 1	2	5.699	5.649	5.749	144		
	3	5.758	5.708	5.808	183		
	4	5.856	5.806	5.906	192		
	5	6.147	6.097	6.197	75.0		
	1	4.915	4.865	4.965	112		
COLUMN 2	2	4.934	4.884	4.984	147	160	
	3	5.112	5.062	5.162	130		
	4	5.154	5.104	5.204	126		
	5	5.381	5.331	5.431	284		
Aroclor-1260	1	7.269	7.219	7.319	149	168	16.08
COLUMN 1	2	7.523	7.473	7.573	208		
	3	7.882	7.832	7.932	151		
	4	8.105	8.055	8.155	162		
	5	8.427	8.377	8.477	170		
	1	6.406	6.356	6.456	151		
COLUMN 2	2	6.593	6.543	6.643	143	143	
	3	6.746	6.696	6.796	132		
	4	7.22	7.17	7.27	151		
	5	7.46	7.41	7.51	139		