

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP070390.D Time Analyzed: 10:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.674	5.581	5.781	481.290	500.000	-3.7
Aroclor-1016-2	5.695	5.602	5.802	489.770	500.000	-2.0
Aroclor-1016-3	5.758	5.665	5.865	477.730	500.000	-4.5
Aroclor-1016-4	5.855	5.762	5.962	481.650	500.000	-3.7
Aroclor-1016-5	6.148	6.056	6.256	472.030	500.000	-5.6
Aroclor-1260-1	7.267	7.175	7.375	489.800	500.000	-2.0
Aroclor-1260-2	7.521	7.429	7.629	480.310	500.000	-3.9
Aroclor-1260-3	7.880	7.788	7.988	516.150	500.000	3.2
Aroclor-1260-4	8.104	8.012	8.212	513.440	500.000	2.7
Aroclor-1260-5	8.425	8.333	8.533	511.510	500.000	2.3
Decachlorobiphenyl	10.247	10.156	10.356	49.160	50.000	-1.7
Tetrachloro-m-xylene	4.521	4.427	4.627	53.520	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP070390.D Time Analyzed: 10:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.914	4.820	5.020	511.470	500.000	2.3
Aroclor-1016-2	4.934	4.839	5.039	512.410	500.000	2.5
Aroclor-1016-3	5.111	5.017	5.217	526.910	500.000	5.4
Aroclor-1016-4	5.152	5.058	5.258	527.990	500.000	5.6
Aroclor-1016-5	5.367	5.273	5.473	527.290	500.000	5.5
Aroclor-1260-1	6.404	6.312	6.512	473.690	500.000	-5.3
Aroclor-1260-2	6.592	6.500	6.700	473.400	500.000	-5.3
Aroclor-1260-3	6.746	6.654	6.854	442.220	500.000	-11.6
Aroclor-1260-4	7.218	7.126	7.326	515.020	500.000	3.0
Aroclor-1260-5	7.460	7.367	7.567	525.520	500.000	5.1
Decachlorobiphenyl	8.878	8.788	8.988	48.040	50.000	-3.9
Tetrachloro-m-xylene	3.827	3.730	3.930	51.950	50.000	3.9