

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP070405.D Time Analyzed: 16:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.679	5.581	5.781	488.210	500.000	-2.4
Aroclor-1016-2	5.700	5.602	5.802	504.420	500.000	0.9
Aroclor-1016-3	5.762	5.665	5.865	495.430	500.000	-0.9
Aroclor-1016-4	5.860	5.762	5.962	506.450	500.000	1.3
Aroclor-1016-5	6.153	6.056	6.256	486.120	500.000	-2.8
Aroclor-1260-1	7.272	7.175	7.375	504.470	500.000	0.9
Aroclor-1260-2	7.526	7.429	7.629	487.930	500.000	-2.4
Aroclor-1260-3	7.884	7.788	7.988	517.560	500.000	3.5
Aroclor-1260-4	8.109	8.012	8.212	503.790	500.000	0.8
Aroclor-1260-5	8.430	8.333	8.533	515.840	500.000	3.2
Decachlorobiphenyl	10.251	10.156	10.356	49.920	50.000	-0.2
Tetrachloro-m-xylene	4.526	4.427	4.627	53.440	50.000	6.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP070405.D Time Analyzed: 16:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.916	4.820	5.020	513.420	500.000	2.7
Aroclor-1016-2	4.935	4.839	5.039	522.180	500.000	4.4
Aroclor-1016-3	5.112	5.017	5.217	542.360	500.000	8.5
Aroclor-1016-4	5.154	5.058	5.258	534.590	500.000	6.9
Aroclor-1016-5	5.369	5.273	5.473	557.910	500.000	11.6
Aroclor-1260-1	6.407	6.312	6.512	494.030	500.000	-1.2
Aroclor-1260-2	6.594	6.500	6.700	498.340	500.000	-0.3
Aroclor-1260-3	6.748	6.654	6.854	466.270	500.000	-6.7
Aroclor-1260-4	7.220	7.126	7.326	532.470	500.000	6.5
Aroclor-1260-5	7.461	7.367	7.567	551.860	500.000	10.4
Decachlorobiphenyl	8.880	8.788	8.988	48.620	50.000	-2.8
Tetrachloro-m-xylene	3.829	3.730	3.930	53.060	50.000	6.1