

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 02/21/2025 02/21/2025

Continuing Calib Time: 09:27 Initial Calibration Time(s): 19:56 21:32

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.16       | 7.16      | 7.06      | 7.26 | 0.00       |
| 2,4-D            | 8.27       | 8.27      | 8.17      | 8.37 | 0.00       |
| 2,4,5-TP(Silvex) | 9.13       | 9.14      | 9.04      | 9.24 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ENTA05

**Lab Code:** CHEM      **Case No.:** Q1523      **SAS No.:** Q1523      **SDG NO.:** Q1523

**Continuing Calib Date:** 03/12/2025      **Initial Calibration Date(s):** 02/21/2025      02/21/2025

**Continuing Calib Time:** 09:27      **Initial Calibration Time(s):** 19:56      21:32

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.62       | 7.63      | 7.53      | 7.73 | 0.01       |
| 2,4-D            | 8.85       | 8.86      | 8.76      | 8.96 | 0.01       |
| 2,4,5-TP(Silvex) | 9.74       | 9.76      | 9.66      | 9.86 | 0.02       |